

CERTIFICATION OF CALIBRATION

ISSUED BY: QED Environmental Systems, Inc. Services Facility

Date Of Calibration: January 28, 2020

Certificate Number: IN12591_11/38402


Page 1 of 1

Approved by Signatory

VIASENSOR

A Division of LANDTEC North America Inc.

QED Environmental Systems, Inc. Services Facility,
2355 Bishop Circle West, Dexter, MI48130
www.qedenv.com


Timothy Hutchins
Laboratory Inspection

Customer: FISHER SCIENTIFIC- PITTSBURGH, PA

PO BOX 1768
PITTSBURGH, PA 15230
USA

Description: Gas Analyser

Model: G100

Serial Number: IN12591

Carbon Dioxide (CO ₂)	
Certified Gas (%)	Instrument Reading (%)
0.0	0.0
2.5	2.4
5.0	4.8
9.9	9.8

Oxygen (O ₂)	
Certified Gas (%)	Instrument Reading (%)
0.0	0.0
5.1	5.1
20.9	20.9

Barometer (mbar)	
Reference	Instrument Reading
0986 mbar / 29.11 "Hg	0983 mbar / 29.03 "Hg

CO₂ readings recorded at: 30.9 °C/87.7 °F

O₂ reading recorded at: 22.8 °C/73.1 °F

Barometric Pressure: 29.11 "Hg

Method of Test: The analyzer is calibrated in a temperature controlled chamber using reference gases.

Calibration Passed: Reported calibration readings are within the manufacturer's specification.

End of Certificate